


**Determination of  $p(\text{CO}_2)$  in air that is  
in equilibrium with a continuous  
stream of sea water**

# Methodology

- Non-dispersive infra-red (NDIR) analyzer  
e.g. LiCOR CO<sub>2</sub>/H<sub>2</sub>O analyzer  
CO<sub>2</sub> gas measurements
- Air-water CO<sub>2</sub> equilibration mechanism  
(flowing SW equilibrates with gas/air)  
Rain-type: showerhead + headspace air  
Bubbler type (bubble seawater with carrier gas)  
Thin film (CO<sub>2</sub> gas permeable membrane)
- Calibration against standard CO<sub>2</sub> gases
- Automation (engineering)



Need correction of H<sub>2</sub>O or removal of H<sub>2</sub>O before detection




Figure 1-1. Schematic of LI-7000 optical path.

Table 1

Summary of main features of the underway  $f\text{CO}_2$  systems “A” through “G” which participated in the exercise

	“A”	“B”	“C”	“D”	“E”	“F”	“G”
<i>Equilibrator</i>							
Design	Shower head	Bubbler	Shower head	Thin film <sup>a</sup>	Showerhead	Bubbler	Showerhead
Total volume	1000 ml	1400 ml	13.1 l	119 ml	11.0 l	36 ml	1200 ml
Water volume	500 ml	1000 ml	2.3 l	21 ml	10.0 l	18 ml	~ 75 ml
Air volume	500 ml	400 ml	10.8 l	98 ml	1.0 l	18 ml	500 ml
Water flow rate	4–6 l min <sup>-1</sup>	2.0 l min <sup>-1</sup>	8.0 l min <sup>-1</sup>	2.0 l min <sup>-1</sup>	10–15 l min <sup>-1</sup>	0 l min <sup>-1b</sup>	1.2 l min <sup>-1</sup>
Air flow rate	0.2 l min <sup>-1</sup>	0.8 l min <sup>-1</sup>	0.5 l min <sup>-1</sup>	1.0 l min <sup>-1</sup>	0.5 l min <sup>-1</sup>	0.17 l min <sup>-1</sup>	0.18 l min <sup>-1</sup>
Vented?	Yes	Yes	Yes	No <sup>c</sup>	Yes	No	Yes
Time constant <sup>d</sup>	2–3 min	75 s	3–5 min	2–3 min	60–90 s	n/a	Unknown
Mean temperature difference <sup>e</sup>	0.30 ± 0.05	0.30 ± 0.04	0.24 ± 0.02	0.39 ± 0.12	0.17 ± 0.03	0.32 ± 0.03	0.56 ± 0.09
<i>CO<sub>2</sub> measurement</i>							
Method	NDIR	NDIR	NDIR	NDIR	NDIR	NDIR	NDIR
Wet/dry?	Wet	Wet	Dry	Dry	Dry	Dry	Wet
<i>Analyzer calibration</i>							
Number of standard gases	2	2	2	2	4	2 <sup>f</sup>	2
Zero gas?	No	Yes	No	No	No	Yes	No
<i>Measurement cycle</i>							
Calibration frequency	6–8 h	6 h	6 h	4–6 h	1.5 h	15 min	2 h
Air measurement frequency	6–8 h	1 h	6 h	4–6 h	0.5 h	n/a	7 min
Interrogation interval	6 s	6 s	1 s	10 s	0.1 s	15 min	0.33 s
Averaging interval	(1 <sup>g</sup> ) 3 min	1 min	4 min	5 min	1 min	n/a	1 s
Reporting interval	(1 <sup>g</sup> ) 3 min	1 min	4 min	5 min	~ 13 min	20 min	~ 8 min
Data points per average	10; 30	10	240	33	600	1	3

<sup>a</sup>Film thickness approximately 0.75 mm.<sup>b</sup>Semi-continuous technique.<sup>c</sup>Vented only every 20 min.<sup>d</sup>This is the overall time constant of the system (not the time constant of equilibration).<sup>e</sup>Mean difference between equilibrator temperature and in situ temperature based on corrected temperature readings (Section 2.3.2).<sup>f</sup>Standard gas generator was initially calibrated using all six calibration gases; linearity checks are carried out for every sample with only two calibration gases.<sup>g</sup>Until June 9, 0230 UTC.

(Kortzinger et al., 2000)





RECONNECT TO WATER TRAP

RECONNECT TO WATER TRAP





Fig. 1. Schematic of the plumbing of the underway system. For the three-way gas solenoid valves *NC* stands for normally closed, *NO* for normally open, and *COM* for common.

Wanninkhof and Thoning, 1993






(Feely et al., 1998)

Fig. 1. Schematic diagram of the automated underway  $p\text{CO}_2$  system. Seawater-equilibrated air first passes through a cold trap and then through a magnesium perchlorate drying tube to remove water vapor prior to analysis by the Li-Cor™ (model 6252) infrared analyzer.



# Annexe



**Fig. 1** Schematic showing the layout of the analytical system described here (based on system designed by Craig Neill, Bjerknes Center for Climate Research, Norway).

(2007 CO<sub>2</sub> Handbook)





